

CERTIFICATE OF CALIBRATION

ISSUED BY: MD CALIBRATION SERVICES

DATE OF ISSUE: 08 Feb 2012

CERTIFICATE NUMBER: H12844

MD Calibration Services

Page 1 of 2

APPROVED SIGNATORY

James Smith

Telephone: +44 (0) 1536 515633 Website: www.mdcalibration.co.uk
Facsimile: +44 (0) 1536 515632 E-mail: info@mdcalibration.co.uk

Customer: iDrill+ Ltd
3 - 4 Ironstone Close Brixworth NN6 9PD **Date Calibrated:** 08 Feb 2012

Identification: 6759 **Customer Order No:** IDP2048

Equipment: Bahco Tape Measure 0 to 5 M

Test Conditions: All calibrations were carried out within the range 20°C +/- 2°C

Basis of Test: Calibrated to M D Calibration procedure 4.15. No Specification - Measured Values Reported.

Report: The above tape has been checked by comparison to master calibrated rules. The measured results are reported on page 2 of this certificate.

Traceability: The reported tests were undertaken on equipment detailed below for which traceability to National Standards is maintained by means of certificates issued by a UKAS accredited laboratory.

Serial No	Equipment Description	Certificate No.
SR1	Rule	108108 (UKAS)

Uncertainty Of Measurement: 0.100mm / 1M

CERTIFICATE OF CALIBRATION

DATE OF ISSUE: 08 Feb 2012

CERTIFICATE NUMBER

H12844

Page 2 of 2

Calibration Results

Feature	Limits	Measured Result
Nominal Position		
0 - 1000mm		1000.0 mm
1000 - 2000mm		1000.2 mm
2000 - 3000mm		1000.2 mm
3000 - 4000mm		1000.1 mm
4000 - 5000mm		1000.0 mm

NOTE: No Specification - Measured Values Reported

* Indicates out of limit, disregarding any uncertainty of measurement.

'n/a' Indicates this measurement is not applicable.

'utc' Indicates unable to calibrate.

**** END OF REPORT ****